

PULSE WIDTH EXTENDER

TLP-8012A5



Pulse Width Extender TLP-8012A5

Advanced TLP/HMM/HBM Solutions

1 Features

- Pulse width extender for the high current TLP/HMM test systems TLP-8010A and TLP-8010C
- 6 manual selectable pulse width: 5, 10, 50, 100, 200 and 500 ns
- High pulse output current up to 80 A

2 Description

The pulse width extender TLP-8012A5 is used to extend the pulse width of the TLP-8010A high-current TLP/HMM test system from 5 ns, 10 ns, 50 ns, 100 ns, 200 ns up to 500 ns in 6 steps.

The pulse width can be selected manually on the backside of the TLP-8012A5 using two high voltage port selection cables.



(a) TLP-8012A5 front side



(b) TLP-8012A5 back side

3 Specifications

Figure 1: TLP-8012A5 pulse width extender

Parameter	Symbol	Limit Values			Unit	Remarks
		Min.	Typ.	Max.		
Pulse width (typical)	t_p	5		500	ns	manual selectable 5, 10, 50, 100, 200, 500 ns
Dimensions TLP-8012A5 (W x H x D)	D_{8012A5}	428 (482.6) x 132.5 x 485			mm ³	428 mm body, 482.6 mm rack flange
Weight TLP-8012A5	W_{8012A5}		11.1		kg	excluding accessories

4 Ordering Information

Pos.	Description	Part No.
01	Pulse width extender TLP-8012A5 with 6 manual selectable built-in pulse width: 5/10/50/100/200/500 ns	TLP-8012A5

General

The product data contained in this data-sheet is exclusively intended for technically trained staff. You and your technical departments will have to evaluate the suitability of the product for the intended application and the completeness of the product data with respect to such application. Our products are solely intended to be commercially used internally and should not be sold to consumers. This data-sheet is describing the specifications of our products for which a warranty is being granted by HPPI GmbH. Any such warranty is granted exclusively pursuant the terms and conditions of the respective supply agreement. There will be no guarantee of any kind for the product and its specifications. For further information on technology, specific applications of our product, delivery terms, conditions and prices please contact HPPI:

High Power Pulse Instruments GmbH
 Stadlerstrasse 6A
 D-85540 Haar, Germany
 Phone : +49 (0)89 8780698 - 440
 Fax : +49 (0)89 8780698 - 444
 E-Mail : info@hppi.de

Due to technical requirements our products and/or their application may be harmful. For information please read carefully the manual or contact HPPI. Safety notes in the manual will inform you about possible risks that result from any foreseeable application of our products. Changes of this data-sheet are reserved.